



Risk Assessment: RSI from prolonged periods of incorrect pipette use.

Carried out by: Samantha Thulborn	Date carried out: April 2014	Review Date: April 2015 – Reviewed – No changes made
		Review Date: April 2016

Hazard (Cause and consequence)	Affected Group	Existing controls	Risk	Further Action Required
RSI caused by poor pipetting technique and posture	All users of pipettes within the RM laboratories.	All personnel pipetting within the RM laboratories will have read the pipetting SOP, which outlines good pipetting practice, which should be followed. All laboratory personnel are properly trained in the use of pipettes, which is recorded in their training folder.	Medium/Low	No